

MercoDIA

Mouse Insulin ELISA

Directions for Use

10-1247-01

Reagents for 96 determinations

10-1247-10

Reagents for 10 X 96 determinations

For Research Use Only





Please note that lot-specific
Calibrator concentration is
stated on vial label

Manufactured by

MercoDIA AB

Sylveniusgatan 8A
SE-754 50 Uppsala
Sweden

Explanation of symbols used on labels

 $\Sigma = 96$	Reagents for 96 determinations
	Expiry date
	Store between 2–8°C
	Lot No.

Intended Use

Mercodia Mouse Insulin ELISA provides a method for the quantitative determination of insulin in mouse serum or plasma.

Summary and explanation of the test

Insulin is the principal hormone responsible for the control of glucose metabolism. It is synthesized in the β -cells of the islets of Langerhans as the precursor, proinsulin, which is processed to form C-peptide and insulin. Both are secreted in equimolar amounts into the portal circulation. The mature insulin molecule comprises two polypeptide chains, the A chain and the B chain. The two chains are linked together by two inter-chain disulphide bridges. There is also an intra-chain disulphide bridge in the A chain.

Secretion of insulin is mainly controlled by plasma glucose concentration, and the hormone has a number of important metabolic actions. Its principal function is to control the uptake and utilisation of glucose in peripheral tissues via the glucose transporter. This and other hypo-glycaemic activities, such as the inhibition of hepatic gluconeogenesis and glycogenolysis are counteracted by the hyperglycaemic hormones including glucagon, epinephrine (adrenaline), growth hormone and cortisol.

Principle of the procedure

Mercodia Mouse Insulin ELISA is a solid phase two-site enzyme immunoassay. It is based on the direct sandwich technique in which two monoclonal antibodies are directed against separate antigenic determinants on the insulin molecule. During incubation insulin in the sample reacts with peroxidase-conjugated anti-insulin antibodies and anti-insulin antibodies bound to microplate wells. A simple washing step removes unbound enzyme labeled antibody. The bound conjugate is detected by reaction with 3,3',5,5'-tetramethylbenzidine. The reaction is stopped by adding acid to give a colorimetric endpoint that is read spectrophotometrically.

Warnings and precautions

- For research use only.
- All samples should be handled as if capable of transmitting infections.
- Each well can only be used once.
- The Stop Solution contains <5% Sulphuric acid.
The Stop Solution is labeled:



Danger

H318 – Causes serious eye damage.

H315 – Causes skin irritation.

P280 – Wear protective gloves. Wear eye or face protection.

P264 – Wash hands thoroughly after handling.

P302 + P352 + P362 + P364 – IF ON SKIN: Wash with plenty of soap and water. Take off contaminated clothing and wash it before reuse.

P332 + P313 – If skin irritation occurs: Get medical attention.

P305 + P351 + P338 + P310 – IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing. Immediately call a POISON CENTER or physician.

- The Enzyme Conjugate Buffer, Cal 0, 1, 2, 3, 4, 5 and Wash Buffer contain <0.06% 5-chloro-2-methyl-4-isothiazolin-3-one and 2-methyl-2H-isothiazol-3-one (3:1).

The Enzyme Conjugate Buffer, the Calibrators and Wash Buffer are labeled:



Warning

H317 – May cause an allergic skin reaction.

P280 – Wear protective gloves.

P261 – Avoid breathing vapour.

P272 – Contaminated work clothing should not be allowed out of the workplace.

P302 + P352 – IF ON SKIN: Wash with plenty of soap and water.

P333 + P313 – If skin irritation or rash occurs: Get medical attention.

P501 – Dispose of contents and container in accordance with all local, regional, national and international regulations.

Material required but not provided

- Pipettes with appropriate volumes (repeating pipettes preferred for addition of enzyme conjugate 1X solution, Substrate TMB and Stop Solution)
- Tubes, beakers and cylinders for reagent preparation
- Redistilled water
- Magnetic stirrer
- Vortex mixer
- Microplate reader with 450 nm filter
- Microplate shaker (700–900 cycles per minute, orbital movement)
- Microplate washing device with overflow function (recommended but not required)

Reagents for 1 X 96 kit

Each Mercodia Mouse Insulin ELISA kit (10-1247-01) contains reagents for 96 wells, sufficient for 42 samples and one Calibrator curve in duplicate. For larger series of assays, use pooled reagents from packages bearing identical lot numbers. The expiry date for the complete kit is stated on the outer label. The recommended storage temperature is 2–8°C.

Coated Plate Mouse monoclonal anti-insulin For unused microplate strips, reseal the bag using adhesive tape, use within 2 months.	1 plate	96 wells 8-well strips	Ready for Use
Calibrators 1, 2, 3, 4, 5 Mouse insulin Color coded yellow Concentration stated on vial label	5 vials	1000 µL	Ready for Use
Calibrator 0 Color coded yellow	1 vial	5 mL	Ready for Use
Enzyme Conjugate 11X Peroxidase conjugated mouse monoclonal anti-insulin	1 vial	1.3 mL	Preparation, see below
Enzyme Conjugate Buffer Color coded blue	1 vial	13 mL	Ready for use
Wash Buffer 21X Storage after dilution: 2–8°C for 2 months	1 bottle	50 mL	Dilute with 1000 mL redistilled water to make wash buffer 1X solution.
Substrate TMB Colorless solution <i>Note! Light sensitive!</i>	1 bottle	22 mL	Ready for Use
Stop Solution 0.5 M H ₂ SO ₄	1 vial	7 mL	Ready for Use

Preparation of enzyme conjugate 1X solution

Prepare the needed volume of enzyme conjugate 1X solution by mixing Enzyme Conjugate 11X with Enzyme Conjugate buffer (1+10) according to the table below. When preparing enzyme conjugate 1X solution for the whole plate, pour all of the Enzyme Conjugate Buffer into the Enzyme Conjugate 11X vial. Mix gently.

Number of strips	Enzyme Conjugate 11X	Enzyme Conjugate Buffer
12 strips	1 vial	1 vial
8 strips	800 µL	8.0 mL
4 strips	400 µL	4.0 mL

Storage after dilution: 2–8°C for 2 months.

Reagents for 10 X 96 kit

Each Mercodia Mouse Insulin ELISA kit (10-1247-10) contains reagents for 10 x 96 wells, sufficient for 42 samples and one calibrator curve in duplicate on each plate. For larger series of assays, use pooled reagents from packages bearing identical lot numbers. The expiry date for the complete kit is stated on the outer label. The recommended storage temperature is 2–8°C.

Coated Plate Mouse monoclonal anti-insulin For unused microplate strips, reseal the bag using adhesive tape, store at 2–8°C and use within 2 months.	10 plate	96 wells 8-well strips	Ready for Use
Calibrators 1, 2, 3, 4, 5 Mouse insulin Color coded yellow Concentration stated on vial label	5 vials	1000 µL	Ready for Use
Calibrator 0 Color coded yellow	1 vial	5 mL	Ready for Use
Enzyme Conjugate 11X Peroxidase conjugated mouse monoclonal anti-insulin	1 vial	12 mL	Preparation, see below
Enzyme Conjugate Buffer Color coded blue	1 vial	120 mL	Ready for use
Wash Buffer 21X	2 bottles	200 mL	Preparation, see below
Substrate TMB Colorless solution <i>Note! Light sensitive!</i>	1 bottle	220 mL	Ready for Use
Stop Solution 0.5 M H ₂ SO ₄	1 vial	70 mL	Ready for Use

Preparation of enzyme conjugate 1X solution

Prepare the needed volume of enzyme conjugate 1X solution by dilution of Enzyme Conjugate 11X in Enzyme Conjugate buffer 1+10 according to the table below. Mix gently.

Number of plates	Enzyme Conjugate 11X	Enzyme Conjugate Buffer
10 plates	1 vial	1 vial
5 plates	5.0 mL	50 mL
3 plates	3.6 mL	36 mL
2 plates	2.4 mL	24 mL
1 plate	1.2 mL	12 mL

Storage after dilution: 2–8°C for 2 months.

Preparation of wash buffer 1X solution

Prepare the needed volume of wash buffer 1X solution by dilution of Wash Buffer 21X in redistilled water 1+20 according to the table below. Mix gently.

Number of plates	Wash Buffer 21X	Redistilled water
10 plates	2 bottles	8000 mL
5 plates	180 mL	3600 mL
3 plates	110 mL	2200 mL
2 plates	70 mL	1400 mL
1 plate	35 mL	700 mL

Storage after dilution: 2–8°C for 2 months.

Specimen collection and handling

Serum

Collect blood by venipuncture, allow to clot, and separate the serum by centrifugation. Samples can be stored at 2–8°C up to 24 hours. For longer periods store samples at -20°C. Avoid repeated freezing and thawing.

Plasma

Collect blood by venipuncture into tubes containing heparin or EDTA as anticoagulant, and separate the plasma fraction. Samples can be stored at 2–8°C up to 24 hours. For longer periods store samples at -20°C. Avoid repeated freezing and thawing.

Preparation of samples

No dilution is normally required for serum and plasma samples, however, samples with a concentration above Calibrator 5 should be diluted in Calibrator 0 (or Mercodia Diabetes Sample Buffer, 10-1195-01).

Test procedure

All reagents and samples must be brought to room temperature before use. Prepare a calibrator curve for each assay run. The product has been optimized and validated without plate sealer.

1. Prepare enzyme conjugate 1X solution, wash buffer 1X solution (according to the tables on previous pages) and samples.
2. Prepare sufficient microplate wells to accommodate Calibrators, controls and samples in duplicate.
3. Pipette 10 μ L each of Calibrators, controls and samples into appropriate wells.
4. Add 100 μ L of enzyme conjugate 1X solution into each well.
5. Incubate on a plate shaker (700–900 rpm) for 2 hours at room temperature (18–25°C).
6. Wash 6 times with 700 μ L wash buffer 1X solution per well using an automatic plate washer with overflow-wash function. After final wash, invert and tap the plate firmly against absorbent paper. Do not include soak step in washing procedure.
Or manually,
Discard the reaction volume by inverting the microplate over a sink. Add 350 μ L wash buffer 1X solution to each well. Discard the wash solution, tap firmly several times against absorbent paper to remove excess liquid. Repeat 5 times. Avoid prolonged soaking during washing procedure.
7. Add 200 μ L Substrate TMB into each well.
8. Incubate 15 minutes on the bench at room temperature (18–25°C).
9. Add 50 μ L Stop Solution to each well.
Place the plate on the shaker for approximately 5 seconds to ensure mixing.
10. Read optical density at 450 nm and calculate results.
Read within 30 minutes.

Note! Be extra careful not to contaminate the Substrate TMB with enzyme conjugate solution.

Internal quality control

Commercial controls such as Mercodia Diabetes Antigen Control Rat/Mouse (L, M, H) (10-1220-01) and/or internal serum pools with low, intermediate and high insulin concentrations should routinely be assayed as samples, and results charted from day to day. It is good laboratory practice to record the following data for each assay: kit lot number, preparation dates of kit components, OD values for the blank, Calibrators and controls.

Calculation of results

The concentration of insulin is obtained by computerized data reduction of the absorbance for the Calibrators, except for Calibrator 0, versus the concentration using cubic spline regression.


Example of results

Wells	Identity	A _{450 nm}	Mean conc. µg/L
1A-B	Calibrator 0	0.072/0.073	
1C-D	Calibrator 1*	0.099/0.101	
1E-F	Calibrator 2*	0.136/0.133	
1G-H	Calibrator 3*	0.466/0.479	
2A-B	Calibrator 4*	1.136/1.118	
2C-D	Calibrator 5*	2.901/2.985	
2E-F	Sample 1	0.163/0.170	0.63
2G-H	Sample 2	0.352/0.361	1.2
3A-B	Sample 3	1.468/1.464	3.7

*Concentration stated on vial label.

Example of calibrator curve

A typical calibrator curve is shown here. Do not use this curve to determine actual assay results.



Conversion factor

1 µg corresponds to 174 pmol.

Limitations of the procedure

Performance limitations

Grossly lipemic, icteric or hemolyzed samples do not interfere in the assay. However, hemolysis in serum and plasma samples may result in a degradation of insulin which could give falsely low values and contributes to higher inter assay variation. The degradation is dependent on time, temperature and the hemoglobin concentration. Keep hemolyzed samples cold or on ice to prevent the insulin degradation.

Expected values

Good practice dictates that each laboratory establishes its own expected range of values.

Performance characteristics

Detection limit

Capability of Detection should be seen as part of a method validation, rather than the lowest concentration that can be measured. The detection limit is $\leq 0.2 \mu\text{g/L}$ as determined by the methodology described in ISO11843- Part 4. Concentration of samples with absorbance below Calibrator 1 should not be calculated, instead expressed as less or equal to (\leq) the concentration indicated on the vial for Calibrator 1.

Recovery

Recovery upon addition is 100%–130% (113%).

Recovery upon dilution is 109%–139% (126%).

Hook effect

Samples with a concentration up to at least 450 $\mu\text{g/L}$ can be measured without giving falsely low results.

Precision

Each sample was analyzed in 4 replicates on 16 different occasions.

Sample	Mean value $\mu\text{g/L}$	Coefficient of variation	
		Repeatability %*	Within laboratory %**
1	0.65	3.1	6.1
2	1.3	1.9	3.5
3	3.6	2.9	5.3

*Within assay variation

**Total assay variation

Specificity

The following cross reactions have been found:

	Crossreaction
IGF-I	n.d.
IGF-II	n.d.
Mouse C-peptide I	n.d.
Mouse C-peptide II	n.d.
Rat C-peptide I	n.d.
Rat C-peptide II	n.d.
Rat insulin	146%
Mouse proinsulin I	43%
Mouse proinsulin II	60%
Rat proinsulin I	14%
Rat proinsulin II	60%
Ovine insulin	256%
Bovine insulin	110%
Human insulin	195%
Human proinsulin	82%
Human C-peptide	n.d.

n.d. = not detected

Calibration

Mercodia Mouse Insulin ELISA is calibrated against an in house reference preparation of mouse insulin.

Warranty

The performance data presented here was obtained using the procedure indicated. Any change or modification in the procedure not recommended by Merckodia AB may affect the results, in which event Merckodia AB disclaims all warranties expressed, implied or statutory, including the implied warranty of merchantability and fitness for use.

Merckodia AB and its authorised distributors, in such event, shall not be liable for damages indirect or consequential.

References

Blyszczuk P, Czyz J, Kania G, Wagner M, Roll U, St-Onge L, Wobus AM. (2003). Expression of Pax4 in embryonic stem cells promotes differentiation of nestin-positive progenitor and insulin-producing cells. *Proc Natl Acad Sci U S A* 100:998-1003

Burcelin R, Crivelli V, Perrin C, Da Costa A, Mu J, Kahn BB, Birnbaum MJ, Kahn CR, Vollenweider P, Thorens B. (2003). GLUT4, AMP kinase, but not the insulin receptor, are required for hepatoportal glucose sensor-stimulated muscle glucose utilization. *J Clin Invest* 111:1555-1562

Friesen NT, Buchau AS, Schott-Ohly P, Lgssiar A, Gleichmann H. (2004). Generation of hydrogen peroxide and failure of antioxidative responses in pancreatic islets of male C57BL/6 mice are associated with diabetes induced by multiple low doses of streptozotocin. *Diabetologia* 47:676-685

Jaekel E, Lipes MA, von Boehmer H. (2004). Recessive tolerance to preproinsulin 2 reduces but does not abolish type 1 diabetes. *Nat Immunol* 5:1028-1035

Further references can be found on our website: www.mercodia.com

Summary of protocol sheet
Mercodia Mouse Insulin ELISA

Add Calibrators, controls* and samples	10 μ L
Add enzyme conjugate 1X solution to all wells	100 μ L
Incubate	2 hours at 18–25°C on a plate shaker, 700–900 rpm
Wash plate with wash buffer 1X solution	700 μ L, 6 times
Add Substrate TMB	200 μ L
Incubate	15 minutes at 18–25°C
Add Stop Solution	50 μ L Shake for 5 seconds to ensure mixing
Measure $A_{450\text{ nm}}$	Evaluate results

*not provided

For full details see page 8

For technical support please contact: support@merck.com